

AvroTros (EenVandaag)

Witte Kruislaan 55
1217 AM Hilversum

19-5-2021

Test Specimen: Specimen: Colour: Client reference number:
A. - 4 Aubergines
B. - 16 Aubergines
C. - 28 Aubergines

Examination: Test number: Test Name:
1. Pesticide testing on food.

Results See following pages

Laboratory Quality Control

Test : 1. Pesticide testing on food. (performed under third party accreditation)
Method : LA-Pesticide-001_07_10/15/2020 based on DIN EN 15662 L00.00-115 §64 LFGB, GC-MS/MS, LC-MS/MS
Measurement uncertainty : SANTE/11813/2017 or lab internal is 50% (Pesticide screening)
Method : LA-GC-022.071_10/22/2018 based on §64 L00.00-36/2 QS-MS extraction and derivation (Bromide)
Product type : Food
Sample taken by : Client.

Results test 1 specimen: A.				
Method: GC-MS/MS Screening	CAS-No.	Amount	Unit	RL
Chlorpyrifos ethyl	2921-88-2	<0.01	mg/kg	0.01
Method: LC-MS/MS Screening				
Acetamiprid	135410-20-7	<0.01	mg/kg	0.01
Carbendazim	1563-66-2	<0.01	mg/kg	0.01
3-OH Carbendazim	16655-82-6	<0.01	mg/kg	0.01
Methomyl	16782-77-5	<0.01	mg/kg	0.01
Dimethoate/omethoate (sum)	-	<0.01	mg/kg	0.01
Method: GC-MS				
Bromide	-	<10	mg/kg	10
RL:	Reporting Limit			
n.d.:	Not detected			

Results test 1 specimen: B.				
Method: GC-MS/MS Screening	CAS-No.	Amount	Unit	RL
Chlorpyrifos ethyl	2921-88-2	<0.01	mg/kg	0.01
Method: LC-MS/MS Screening				
Acetamiprid	135410-20-7	<0.01	mg/kg	0.01
Carbendazim	1563-66-2	<0.01	mg/kg	0.01
3-OH Carbendazim	16655-82-6	<0.01	mg/kg	0.01
Methomyl	16782-77-5	<0.01	mg/kg	0.01
Dimethoate/omethoate (sum)	-	<0.01	mg/kg	0.01
Method: GC-MS				
Bromide	-	<10	mg/kg	10
RL:	Reporting Limit			
n.d.:	Not detected			

Results test 1 specimen: C.				
Method: GC-MS/MS Screening	CAS-No.	Amount	Unit	RL
Chlorpyrifos ethyl	2921-88-2	<0.01	mg/kg	0.01
Method: LC-MS/MS Screening				
Acetamiprid	135410-20-7	<0.01	mg/kg	0.01
Carbendazim	1563-66-2	<0.01	mg/kg	0.01
3-OH Carbendazim	16655-82-6	<0.01	mg/kg	0.01
Methomyl	16782-77-5	<0.01	mg/kg	0.01
Dimethoate/omethoate (sum)	-	<0.01	mg/kg	0.01
Method: GC-MS				
Bromide	-	<10	mg/kg	10
RL:	Reporting Limit			
n.d.:	Not detected			

Laboratory Quality Control